



# Certification of Conformity

EU-Medical Devices Directive-93/42/EEC

Registration No.: ENC120510GZ20-1

Applicant : Shenghua Beauty Equipment Co., LTD

Applicant Address : No.485, Zhonghua North Street, Handan City 056004, Hebei Province, China

Product Designation : Derma roller system

Model Number : BIO50,SH501,SH540,SH541,SH531,SH192,SH500  
DermaVib,ULTIMA-A3,ULTIMA N2,ULTIMA A1,M5,M7  
Needle Cartridge,DRS1200,DRS1080,Derma roller 192,DRS180

Manufacturer : Shenghua Beauty Equipment Co., LTD

Manufacturer Address : No.485, Zhonghua North Street, Handan City 056004, Hebei Province, China

The submitted products have been tested by us with the listed standard and found in compliance with the Medical Devices Directive 93/42/EEC and following European Standards:

| Directive | Applied Standards                | Test Report No. |
|-----------|----------------------------------|-----------------|
| Safety    | EN 60601-1:2006+A11:2011+A1:2013 | ENC180411GZ82L1 |
| EMC       | EN60602 -1-2:2015                | ENC180411GZ82E1 |
|           | EN 61000-3-2:2014                |                 |
|           | EN 61000-3-3:2013                |                 |

'This certificate is based on an evaluation of a sample of the above mentioned product. Technical report and documentation are at the licence applicant's disposal. This certificate does not imply assessment of the series-production of the product. The CE markings as shown below can be affixed on the product after preparation of necessary technical documentation.



Ray Zhou / General Manager

Issue Date: May 10, 2018

**East Notice Certification Service Co., Ltd.**

317/319 3/F, HuiCheng Building, XiangPu Street, DongPu, TianHe, GuangZhou

Tel: +86-020-2331 4234 Fax: +86-020-8256 8534 E-mail: enc@enc-lab.com Web: www.enc-lab.com